

GOVT. OF ASSAM
OFFICE OF THE PRINCIPAL CUM CHIEF SUPERINTENDENT
JORHAT MEDICAL COLLEGE & HOSPITAL, JORHAT
(Under Society for Medical Education Jorhat)

No. SMEJ/JMCH/2235/2018/

Dated Jorhat, the 25th July, 2018

SHORT QUOTATION NOTICE

Short sealed quotations affixing court fee stamp of Rs. 8.25 (Rupees Eight and Paise Twenty Five) only are invited from intending Suppliers/ Traders and Manufacturer for supply of ETO Sterilizer (Annexure-I) for Jorhat Medical College & Hospital, Jorhat.

The sealed quotations are to be submitted to the undersigned on or before 11:00 AM of 04.08.2018 and shall be opened on the same day at 12:00 Noon in presence of the questioners or their authorized representatives. Rates of each items are to be quoted separately.

Bidders are also required to submit a non-refundable amount of Rs. 1000/- (One thousand) only by way of Demand Draft to be drawn in favour of 'Jorhat Medical College Hospital Users' Fund, Jorhat' as 'Cost of Document'.

The bidders are requested to comply to the newly implemented GST guidelines of GOI.

Terms & Conditions:

1. Demand Draft for Rs. 1000/- only to be drawn in favour of 'Jorhat Medical College Hospital Users' Fund, Jorhat.'
2. A sum of Rs. 10,000/- (Rupees Ten Thousand) shall have to be submitted by way of Demand Draft favouring 'Jorhat Medical College Hospital Users' Fund, Jorhat' as Earnest Money along with the Quotation
3. Financial Ability Certificate from component authority along with minimum turnover of Rs. 10.00 lakhs per annum.
4. Tenderer's must be having prior working experience of not less than three years in Govt./PSO/Private Organization. They shall have to submit Performance Certificate from concerned organization/ Establishment.
5. The tenderer must have executed similar works before and should have minimum supply orders from Govt./PSO/Pvt. Organization of Rs. 10.00 lakhs in the last three years.
6. Registration Certificate of the Firm under Society Act; in case of Society
7. Photocopy of PAN Card/GRN No./TIN No.
8. GST Registration Certificate.
9. All quotations must accompany up-to-date Income Tax and Sales Tax Clearance certificates.
10. All the papers of the quotation must be signed with date.
11. The undersigned reserves the right to accepts or reject any or all the quotations without showing any reasons thereof.



Principal cum Chief Superintendent
Jorhat Medical College & Hospital, Jorhat

Dated Jorhat, the 25th July, 2018

Memo No. SMEJ/JMCH/2235/2018/


Copy to:

- 1) The Director of Medical Education, Assam, Sixth Mile, Khanapara, Guwahati-22.
- 2) The Director, Information and Public Relations, Assam, Dispur, Guwahati – 6 with a request for taking up necessary steps of publishing the advertisement in local English and Assamese daily newspapers published from Guwahati and Jorhat in one issue.
- 3) The Deputy Commissioner, Jorhat.
- 4) The Superintendent, JMCH, Jorhat.
- 5) The Member Secretary, Tender Committee, JMCH, Jorhat
- 6) Notice Boards of -
 - a) The Deputy Commissioner's Office, Jorhat.
 - b) The Treasury Office, Jorhat.
 - c) The Office of the Superintendent, Jorhat Medical College & Hospital, Jorhat.
- 7) Office copy / Website.


Principal cum Chief Superintendent
Jorhat Medical College & Hospital, Jorhat

TECHNICAL SPECIFICATION OF STERILIZATION EQUIPMENT

Sl. No	Name of Item	Specification and Scope of works
1	Ethylene Oxide (ETO) Sterilizer	<ol style="list-style-type: none"> 1) Capacity Should Be Around 50 Litres. 2) Power supply should be 220v AC/50Hz. 3) The Chamber Should Be Double Walled Made of Suitable Material Which Is Resistant to Corrosion and Gas. 4) To Minimize the Gas Deposit the Interior of The Chamber Should Be Smoothly Finished 5) There Should Be Provision for Heating the Chamber Either with Strip Air Heater or Hot Water Circulated Through a Coil of Stainless Steel Around the Chamber to Maintain the Chamber Temperature Around 30 – 55 Degrees Centigrade. 6) The Chamber Shall Be Appropriately Insulated Against Heat Emission. 7) The Door of The Steriliser Should Have a Quick Release Locking Arrangement, With Door Opening to The Sides 8) Suitable safety interlocking arrangement shall be provided for the door so that the sterilization process does not start unless the door is properly locked in position and should not open during the programme. 9) The Sterilizer shall be provided with a compatible pump which should be able to Separate and Evacuate the Gas and the exhaust line should be above 12 Mtr. of building height. 10) Compatible pump should be able to achieve elevated level of Air Removal and should be noise and maintenance free. 11) The Sterilization Process Should Be Fully Automatic. With Two Numbers Sterilization Cycles with Inbuilt Leak Test and Aeration Should Be Provided, Which Should Be Completed Automatically. 12) The Stand Should Be Made of Appropriate material. 13) Documentation – Thermal Printer / Dot-Matrix Printer Which Should Print Date, Batch/Load Number, Program Type Elected and Program Parameters Which Includes One Point Pressure and One Point Temperature Print Out. 14) The ETO Sterilizer Shall Be Equipped with The Following Accessories: <ol style="list-style-type: none"> a) Sterilization Basket of Suitable Size: 1 No. b) Eon Gas Cartridges: Approximately 100 No. And Should Be Pea Certified c) Packaging Material with Chemical Indicator of All Sizes :10 rolls. 15) The ETO Sterilizer Should Have Compliance to ISO 13485:1996. Certificate to Be Provided by Equipment Manufacturer. (or equivalent) 16) Product should be Indian make. 18) The Consumable Required For 200 Cycles to Be Provided. 19) Warranty : 1 years <p>Scope of Work:</p> <ol style="list-style-type: none"> i) The supplier has to undertake the supply, installation, testing and commissioning of the ETO sterilizers ii) These works shall comply with all relevant safety and standard guidelines. iii) The supplier shall obtain all necessary clearances for the commissioning of the ETO sterilizers in liaison with the consignee


 Principal cum Chief Superintendent,
 Jorhat Medical College & Hospital, Jorhat